

# 普源精电科技股份有限公司

## 计量实验室

RIGOL Technologies CO., LTD. Metrology Lab

# 校准证书

CALIBRATION CERTIFICATE

客户名称:

Customer

普源精电科技股份有限公司

客户地址:

Address of customer

苏州市高新区科灵路8号

样品名称:

Name of Sample

机架示波器

型号规格:

Model /Type

DS8034-R

制造单位:

Manufacturer

RIGOL

样品编号:

No. of sample

DS8T231100013

校准日期:

Date of Calibration

2021-03-17

接收日期:

Receive Date

2021-03-16

校准人员:

Calibrated by

王燕

核验人员:

Checked by

许昌发

批准人员:

Approved by

志琦

职务:

Position

质量监督员

签发日期:

Date of Calibration

2021-03-18



地址: 苏州市高新区科灵路8号

Address No.8 Ke Ling Road, New District,Suzhou, China

邮编: 215163

PostCode

电话: 0512-66706688-22067

Tel.

网址: [www.rigol.com](http://www.rigol.com)

Web



中国认可  
国际互认  
校准  
CALIBRATION  
CNAS L11178

证书编号: RSZCAL2103001  
certificate No.

本实验室质量管理体系符合 ISO/IEC 17025 的要求。

This laboratory quality management system meets the ISO/IEC17025.

本次校准所依据的方法规范:

Reference documents for the calibration methods

JJG 262-1996 《模拟示波器检定规程》 JJF 1057-1998 《数字存储示波器校准规范》

本次校准结果判定所依据的技术规范:

Technical specifications for the determination of the results of calibration

《产品数据手册》

本次校准使用的主要测量标准:

Main Standards of Measurement Used in the Calibration

名称 Name	型号规格 Model	编号 NO.	测量范围 Measure Range	不确定度或准确度等级或最大允许误差 Uncertainty or Accuracy Class or MPE	溯源机构 Traceability Mechanism	证书编号/有效期 Certificate No./ Due Date
示波器校准仪	9500B/ 9530	9500B 主机: 936954411 9530 探头: 151861813	直流电压 (1MΩ):±(1mV~200V), (50Ω):±(1mV~5V); 方波电压 (1MΩ):5mVpp~200Vpp; (50Ω):5 mVpp~5Vpp; 稳幅正弦波:5mVpp~5Vpp (0.1Hz~1.1GHz); 时标信号:0.2ns~55s; 快沿脉冲: 150ps	MPE: 直流电压: ±(0.025%+25uV); 方波电压 (<10kHz): ±(0.1%+10uV); 稳幅正弦波平坦度: ±5% ; 时标信号:±0.25ppm; 快沿脉冲:±25ps	苏州市计量测试院	801500826/ 2021.10.20

校准地点和环境条件:

Location and environmental condition for the calibration

地点: 校准室

Location

温度: 21.3 °C

湿度: 41.4 %RH

Temperature

Humidity

- 注: 1、本证书的校准结果仅对当时所校准样品有效。The results of this certificate are only responsible for the item calibrated.
- Statement 2、未经本实验室书面批准,不得部分复制。This certificate shall not be reproduced except in full, without the written approval of our laboratory.
- 3、本证书未加盖实验室印章无效。This certificate is invalid without official stamp
- 4、证书结果中“P”代表“合格”,“F”代表“不合格”,“N/A”代表“不适用”。In the results of the certification,“P”stands for “Pass”,“F”stands for “Fail”,and “N/A”stands for “No applicable” .
- 5、除特殊情况,本证书给出的符合性说明未考虑不确定度。Except for special cases, the monformance statement given in this certificate does not consider uncertainty.

## 校准结果 / 说明

### Results of Calibration and additional explanation

1、外观和功能检查: 符合要求

Appearance and Function Check

2、计量特性测量结果:

the results unrelated to metrological specification

2.1 扫描时间系数:

偏转系数 ns/div	下限值 (ppm)	误差	上限值 (ppm)	测量不确定度 $U_{rel} (k=2)$	结论 Pass/Fail
1.000	-3	0.0%	3	0.3%	P
2.000	-3	0.0%	3	0.3%	P
5.000	-3	0.0%	3	0.3%	P
10.00	-3	0.0%	3	0.3%	P
20.00	-3	0.0%	3	0.3%	P
50.00	-3	0.0%	3	0.3%	P
100.0	-3	0.0%	3	0.3%	P
200.0	-3	0.0%	3	0.3%	P
500.0	-3	0.0%	3	0.3%	P
μs/div	/	/	/	/	/
1.000	-3	0.0%	3	0.3%	P
2.000	-3	0.0%	3	0.3%	P
5.000	-3	0.0%	3	0.3%	P
10.00	-3	0.0%	3	0.3%	P
20.00	-3	0.0%	3	0.3%	P
50.00	-3	0.0%	3	0.3%	P
100.0	-3	0.0%	3	0.3%	P
200.0	-3	0.0%	3	0.3%	P
500.0	-3	0.0%	3	0.3%	P
ms/div	/	/	/	/	/
1.000	-3	0.0%	3	0.3%	P
2.000	-3	0.0%	3	0.3%	P
5.000	-3	0.0%	3	0.3%	P
10.00	-3	0.0%	3	0.3%	P
20.00	-3	0.0%	3	0.3%	P
50.00	-3	0.0%	3	0.3%	P
100.0	-3	0.0%	3	0.3%	P
200.0	-3	0.0%	3	0.3%	P
500.0	-3	0.0%	3	0.3%	P

## 校准结果 / 说明

### Results of Calibration and additional explanation

续上表:

偏转系数 s/div	下限值 (ppm)	误差	上限值 (ppm)	测量不确定度 $U_{rel} (k=2)$	结论 Pass/Fail
1.000	-3	0.0%	3	0.3%	P
2.000	-3	0.0%	3	0.3%	P
5.000	-3	0.0%	3	0.3%	P

#### 2.2.1 垂直偏转系数 (1M $\Omega$ ):

标称值 mV/div	下限值 (mV)	实际值 (mV)				上限值 (mV)	测量不确定度 $U_{rel} (k=2)$	结论 Pass/Fail
		CH1	CH2	CH3	CH4			
1.00	5.52	5.92	5.88	5.96	5.81	6.48	0.4%	P
2.00	11.52	11.97	11.89	11.93	11.88	12.48	0.4%	P
5.00	29.4	29.57	29.87	29.58	29.81	30.6	0.4%	P
10.0	58.8	59.65	59.30	59.13	59.32	61.2	0.4%	P
20.0	117.6	118.5	119.9	118.0	119.7	122.4	0.4%	P
50.0	294	297.6	298.3	295.6	298.0	306	0.4%	P
100	588	594.8	595.0	596.1	597.8	612	0.4%	P
/	(V)	(V)				(V)	/	/
200	1.176	1.184	1.192	1.185	1.187	1.224	0.4%	P
500	2.94	2.97	2.96	2.96	2.99	3.06	0.4%	P
V/div	/	/				/	/	/
1.0	5.88	5.96	5.96	5.92	5.97	6.12	0.4%	P
2.0	11.76	11.98	11.86	11.89	11.91	12.24	0.4%	P
5.0	29.4	29.62	29.89	29.84	29.78	30.6	0.4%	P
10.0	58.8	59.36	59.75	59.19	59.36	61.2	0.4%	P

## 校准结果 / 说明

### Results of Calibration and additional explanation

#### 2.2.2 垂直偏转系数 (50Ω):

标称值 mV/div	下限值 (mV)	实际值 (mV)				上限值 (mV)	测量不确定度 $U_{rel} (k=2)$	结论 Pass/Fail
		CH1	CH2	CH3	CH4			
1.00	5.52	6.09	6.17	5.88	5.92	6.48	0.4%	P
2.00	11.52	11.97	11.91	11.99	11.84	12.48	0.4%	P
5.00	29.4	29.82	29.78	30.08	29.62	30.6	0.4%	P
10.0	58.8	59.64	59.25	60.08	59.40	61.2	0.4%	P
20.0	117.6	119.2	119.0	119.4	118.4	122.4	0.4%	P
50.0	294	298.9	297.0	298.1	297.7	306	0.4%	P
100	588	596.0	596.8	593.8	592.0	612	0.4%	P
/	(V)	(V)				(V)	/	/
200	1.176	1.201	1.194	1.190	1.185	1.224	0.4%	P
500	2.94	2.99	2.98	2.97	2.95	3.06	0.4%	P

#### 2.3 频带宽度:

垂直偏转系数 (mV/div)	下限值 (MHz)	实测值 (MHz)				测量不确定度 $U (k=2)$	结论 Pass/Fail
		CH1	CH2	CH3	CH4		
2.00	350	575	575	565	560	5%	P
5.00		580	585	575	572	5%	P
10.0		573	579	575	570	5%	P
20.0		585	579	575	565	5%	P
50.0		585	580	575	575	5%	P
100		590	580	575	573	5%	P
200		594	595	588	586	5%	P

2.4 上升时间:

垂直偏转系数 (mV/div)	实测值(ps)				上限值 (ns)	测量不确定度 $U (k=2)$	结论 Pass/Fail
	CH1	CH2	CH3	CH4			
2.00	610	580	580	600	1.00	4%	P
5.00	560	580	580	580		4%	P
10.0	550	560	560	560		4%	P
20.0	530	540	560	540		4%	P
50.0	530	560	540	560		4%	P
100	510	540	540	560		4%	P
200	500	500	520	520		4%	P

说明 (Direction):

- 1、经校准，所校项目符合判定依据规定的技术要求。

On the basis of calibration results, it has been found that the instrument calibrated meet metrological.

- 2、建议复校时间间隔为 12 个月。

It's recommended that the sample is recalibrated within 12 months.

- 3、依据《JJF1059.1-2012 测量不确定度评定与表示》。

Reference Document 《JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement》.

以下空白

(Blank Below)